


If in doubt – Ask

Dimensions in Millimeters Unless Otherwise Stated

Do Not Scale This Drawing

Purch Information[ Note (1) ]	
Description	Example
Job Reference	EUJ XXXX
Material	XXXX
Connection	XXXX
Length	XXXX mm
Tag No.	XXXX
Heat No.	HT XXXX
Stem Heat No.	S,HT XXXX
Serial No.	S,N XXXX



- Note (1) : Punch on the ring of the flange
- Note (2) : Full penetration weld requires to have min. pipe sch80.
- Note (3) : Clean internally from welding mark and dusts
- Note (4) : Direction of flow (Pls follow process flow direction)
- Note (5) : Hydrostatic Test Pressure will follow ASME Flange Rating or Valve Design Pressure, whichever is lower.
- Note (6) : Number of bolt holes will be depend on flange rating

5	Tee Connector								
4	Nipple								
3	Gate Valve								
2	Flange								
1	Seamless Pipe								
S/N	DESCRIPTION	A.S PART NO.	MATERIAL	QTY.	REMARK	DRAWN BY	APPROVED BY	STD DWG NO.	ASSP-01-14

Fixed Type-1 Sample Probe

Tolerance Length < 1 mtr (±1mm), 1 mtr < Length (±2mm) < 2 mtr

CUSTOMER

Wai Lin

APPROVED BY

ACEZ SENSING PTE LTD

28E PENJURU CLOSE, #01-04

SINGAPORE 609133

TEL: 65-6265 1588 FAX : 65-6264 3992

Website : www.acezsensing.com